

	SSK (Six revision procedures)	OSS (Nine revision procedures)
Mean arc of motion at presentation	54 degrees	45 degrees
Mean gain in arc of motion	39 degrees	55 degrees
Mean flexion contracture at presentation	28 degrees	26 degrees
Mean correction of flexion contracture	23 degrees	24 degrees

Table 4 (Comparison of Constrained Condylar and Rotating-hinge implants) SSK: Super-Stabilized Knee (Biomet Inc., Warsaw, Indiana), OSS: Orthopedic Salvage System (Biomet Inc., Warsaw, Indiana)